
Midmark M11 - v1.0.0
Total Cycles: 2883

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2023 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 222.9 F 0.0
2:00 215.1 F 4.6
4:00 223.6 F 10.8
6:00 229.8 F 6.7
8:00 240.7 F 12.7
10:00 240.6 F 11.0
12:00 239.7 F 10.1
14:00 254.1 F 17.3
16:00 266.0 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 28.9
2:00 271.5 F 28.2
3:00 272.3 F 29.1
4:00 272.1 F 28.7

Min 271.1 F 28.1
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 237.2 F 0.0
10:00 230.8 F 0.0
15:00 228.6 F 0.0
20:00 228.3 F 0.0
25:00 228.5 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 2:42
HEATING: 17:38
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

02/09/2023 03:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____