
Midmark M11 - v1.0.0
Total Cycles: 1977

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2022 06:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.1 F 0.0
2:00 75.9 F 0.0
4:00 81.0 F 0.1
6:00 154.9 F 0.2
8:00 144.0 F 2.3
10:00 168.3 F 6.1
12:00 188.2 F 10.0
14:00 204.7 F 9.1
16:00 221.5 F 7.0
18:00 232.7 F 10.8
20:00 233.0 F 7.7
22:00 240.4 F 11.4
24:00 235.7 F 8.5
26:00 249.0 F 14.8
28:00 259.7 F 20.8
30:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.4 F 29.0
2:00 272.4 F 29.4
3:00 271.2 F 29.0
4:00 272.0 F 28.9

Min 271.1 F 28.7
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 212.2 F 0.0
10:00 204.1 F 0.0
15:00 198.4 F 0.0
20:00 191.7 F 0.0
25:00 196.4 F 0.0
30:00 193.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:40
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

05/31/2022 08:43 AM
ACCEPTED: X
REJECTED:
Operator: U02