
Midmark M11 - v1.0.0
Total Cycles: 4879

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2024 12:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.4 F 0.0
2:00 126.4 F 0.0
4:00 137.1 F 0.3
6:00 170.2 F 0.7
8:00 194.8 F 3.8
10:00 209.8 F 8.0
12:00 223.7 F 12.2
14:00 227.4 F 7.5
16:00 240.6 F 12.7
18:00 239.3 F 10.4
20:00 235.9 F 8.6
22:00 250.2 F 15.4
24:00 261.8 F 22.0
26:00 272.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.3 F 29.2
2:00 272.5 F 29.1
3:00 271.7 F 28.5
4:00 272.6 F 29.4

Min 271.6 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 204.0 F 0.0
10:00 219.4 F 0.0
15:00 230.6 F 0.0
20:00 228.4 F 0.0
25:00 228.1 F 0.0
30:00 228.6 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 27:06
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:05:27
CYCLE COMPLETE

09/09/2024 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____