
Midmark M11 - v1.0.0
Total Cycles: 5293

Name: RITTER 1
Sterilizer ID: V2178198
01/29/2025 02:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.1 F 0.0
2:00 157.9 F 3.3
4:00 175.6 F 7.1
6:00 194.7 F 11.0
8:00 220.8 F 3.9
10:00 223.6 F 7.4
12:00 235.2 F 11.9
14:00 233.8 F 8.5
16:00 243.5 F 12.0
18:00 244.0 F 12.5
20:00 255.4 F 18.5
22:00 266.0 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.2 F 28.7
2:00 272.1 F 29.4
3:00 273.0 F 29.8
4:00 272.8 F 29.6

Min 271.0 F 28.4
Max 273.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 225.2 F 0.0
10:00 216.9 F 0.0
15:00 220.1 F 0.0
20:00 226.0 F 0.0
25:00 230.4 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:43
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:02:42
CYCLE COMPLETE

01/29/2025 03:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____