
Midmark M11 - v1.0.5
Total Cycles: 4058

Name: RITTER2
Sterilizer ID: V2313228
10/25/2024 03:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.8 F 0.0
2:00 171.7 F 3.3
4:00 197.0 F 8.6
6:00 214.9 F 8.4
8:00 228.3 F 9.8
10:00 233.5 F 8.6
12:00 242.8 F 12.0
14:00 246.1 F 13.6
16:00 259.3 F 20.9
18:00 271.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.3 F 29.3
2:00 272.2 F 29.6
3:00 271.5 F 28.8
4:00 272.3 F 29.3

Min 271.4 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 224.2 F 0.0
10:00 214.4 F 0.0
15:00 199.2 F 0.0
20:00 193.9 F 0.0
25:00 193.4 F 0.0
30:00 194.7 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 18:31
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

10/25/2024 04:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____