
Midmark M11 - v1.0.0
Total Cycles: 2242

Name: RITTER 1
Sterilizer ID: V2178198
08/24/2022 04:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 148.0 F 0.1
4:00 169.0 F 0.7
6:00 195.4 F 3.9
8:00 216.1 F 8.3
10:00 229.8 F 13.3
12:00 232.9 F 9.4
14:00 236.6 F 9.4
16:00 244.8 F 13.1
18:00 243.6 F 12.0
20:00 256.8 F 18.8
22:00 268.2 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.8
2:00 271.6 F 28.4
3:00 272.5 F 28.8
4:00 271.5 F 28.2

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 228.5 F 0.0
10:00 238.0 F 0.0
15:00 231.8 F 0.0
20:00 224.6 F 0.0
25:00 226.0 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:14
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

08/24/2022 05:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____