
Midmark M11 - v1.0.5
Total Cycles: 1396

Name: RITTER2
Sterilizer ID: V2313228
08/16/2022 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 192.4 F 1.6
4:00 196.9 F 7.1
6:00 214.5 F 12.8
8:00 228.3 F 8.7
10:00 239.1 F 11.1
12:00 241.8 F 12.5
14:00 241.4 F 11.6
16:00 254.4 F 18.3
18:00 265.8 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.6 F 29.6
2:00 271.8 F 29.2
3:00 272.7 F 29.7
4:00 271.8 F 29.0

Min 271.7 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 205.2 F 0.0
10:00 219.8 F 0.0
15:00 223.9 F 0.0
20:00 223.5 F 0.0
25:00 221.9 F 0.0
30:00 220.4 F 0.0

Drying Complete

FILLING: 2:10
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

08/16/2022 10:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____