
Midmark M11 - v1.0.0
Total Cycles: 4666

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2024 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 85.6 F 0.0
4:00 105.8 F 0.1
6:00 116.0 F 0.3
8:00 142.3 F 1.5
10:00 169.7 F 4.3
12:00 196.2 F 7.2
14:00 206.8 F 11.2
16:00 213.9 F 4.0
18:00 224.4 F 7.1
20:00 235.1 F 11.1
22:00 231.6 F 7.1
24:00 242.8 F 12.3
26:00 239.9 F 10.1
28:00 252.3 F 16.2
30:00 263.0 F 22.4
32:00 272.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 272.4 F 28.8
2:00 271.7 F 28.0
3:00 272.6 F 28.8
4:00 272.0 F 28.2

Min 271.3 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 230.5 F 0.0
10:00 225.4 F 0.0
15:00 230.3 F 0.0
20:00 226.5 F 0.0
25:00 224.2 F 0.0
30:00 224.4 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:12
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:22
CYCLE COMPLETE

06/21/2024 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____