
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000688
9:39 15/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:12
STERILIZING 3:15
136.0 C 229kPa 3:15
136.4 C 231kPa 3:45
136.4 C 232kPa 4:15
136.4 C 233kPa 4:45
136.3 C 231kPa 5:15
136.2 C 232kPa 5:45
136.4 C 232kPa 6:15
136.3 C 233kPa 6:45
136.3 C 233kPa 6:45
VENTING 6:45
AIR DRYING 7:37
CYCLE COMPLETE 7:41

Digital Signature #
CEC1383E30F22160
