

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001173  
16:26 18/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:23  
PRESSURIZING 2:33  
STERILIZING 3:42  
136.0 C 220kPa 3:42  
136.4 C 223kPa 4:13  
136.1 C 221kPa 4:43  
136.2 C 224kPa 5:13  
136.1 C 222kPa 5:43  
136.1 C 222kPa 6:13  
136.4 C 225kPa 6:43  
136.4 C 225kPa 7:13  
136.4 C 225kPa 7:13  
VENTING 7:13  
AIR DRYING 8:06  
CYCLE COMPLETE 20:39  
-----

Digital Signature #  
88D3CF56E77A5E68  
-----