
Midmark M11 - v1.0.0
Total Cycles: 2310

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2022 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.7 F 0.0
2:00 193.5 F 4.0
4:00 212.6 F 8.2
6:00 229.4 F 13.1
8:00 218.2 F 7.4
10:00 232.1 F 12.7
12:00 234.6 F 9.6
14:00 236.3 F 8.7
16:00 247.5 F 14.2
18:00 259.6 F 20.9
20:00 270.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.1 F 29.0
2:00 272.5 F 29.5
3:00 271.5 F 28.7
4:00 272.4 F 29.3

Min 271.5 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 222.7 F 0.0
10:00 233.0 F 0.0
15:00 229.7 F 0.0
20:00 229.3 F 0.0
25:00 230.6 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 20:46
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

09/12/2022 12:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____