
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002284
7:42 25/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:18
136.5 C 220kPa 4:18
136.4 C 221kPa 4:48
136.3 C 223kPa 5:18
136.3 C 220kPa 5:48
136.4 C 224kPa 6:18
136.2 C 217kPa 6:48
136.4 C 224kPa 7:18
136.0 C 220kPa 7:48
136.0 C 220kPa 7:48
VENTING 7:48
AIR DRYING 8:39
CYCLE COMPLETE 8:45

Digital Signature #
C59D7F5B17E5B7AD
