

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1948

Name: RITTER 1  
Sterilizer ID: V2178198  
05/25/2022 07:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.0 F 0.0  
2:00 75.5 F 0.1  
4:00 85.5 F 1.3  
6:00 114.4 F 3.6  
8:00 150.4 F 6.5  
10:00 174.1 F 9.9  
12:00 192.2 F 12.0  
14:00 208.2 F 4.2  
16:00 219.1 F 7.8  
18:00 229.4 F 11.6  
20:00 229.4 F 6.6  
22:00 240.6 F 11.9  
24:00 237.9 F 9.3  
26:00 250.7 F 15.5  
28:00 261.6 F 21.7  
30:00 271.8 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.5 F 28.9  
2:00 271.4 F 28.0  
3:00 272.5 F 29.0  
4:00 271.8 F 28.2

Min 271.3 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.4 F 0.0  
5:00 215.2 F 0.0  
10:00 215.8 F 0.0  
15:00 229.6 F 0.0  
20:00 240.9 F 0.0  
25:00 229.7 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:36  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:08:29  
CYCLE COMPLETE

05/25/2022 08:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_