
Midmark M11 - v1.0.0
Total Cycles: 1529

Name: RITTER 1
Sterilizer ID: V2178198
01/14/2022 05:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.7 F 0.0
2:00 173.7 F 2.0
4:00 190.5 F 5.6
6:00 204.3 F 9.7
8:00 219.4 F 4.8
10:00 227.2 F 8.1
12:00 239.3 F 13.0
14:00 237.8 F 9.8
16:00 238.9 F 8.9
18:00 248.7 F 14.2
20:00 260.2 F 20.6
22:00 271.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.3 F 28.8
2:00 272.1 F 28.3
3:00 271.0 F 27.9
4:00 272.1 F 29.0

Min 271.0 F 27.9
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.4 F 0.0
5:00 212.2 F 0.0
10:00 213.9 F 0.0
15:00 215.4 F 0.0
20:00 215.4 F 0.0
25:00 216.7 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

01/14/2022 06:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____