
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002587
12:21 21/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:19
STERILIZING 3:20
136.0 C 222kPa 3:20
136.3 C 225kPa 3:50
136.4 C 225kPa 4:20
136.4 C 225kPa 4:50
136.2 C 224kPa 5:20
136.1 C 222kPa 5:50
136.4 C 225kPa 6:20
136.4 C 225kPa 6:50
136.4 C 225kPa 6:50
VENTING 6:50
AIR DRYING 7:41
CYCLE COMPLETE 37:41

Digital Signature #
D4206D593C518B48
