
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006737
14:05 13/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:36
STERILIZING 3:37
136.5 C 223kPa 3:37
136.4 C 224kPa 4:07
136.4 C 223kPa 4:37
136.3 C 220kPa 5:07
136.1 C 221kPa 5:37
136.2 C 220kPa 6:07
136.3 C 222kPa 6:37
136.4 C 223kPa 7:07
136.4 C 223kPa 7:07
VENTING 7:07
AIR DRYING 8:05
CYCLE COMPLETE 8:12

Digital Signature #
82FF02222AA39871
