

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001581  
12:53 09/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:48  
STERILIZING 3:48  
136.0 C 220kPa 3:48  
136.2 C 223kPa 4:18  
136.3 C 224kPa 4:48  
136.3 C 222kPa 5:18  
136.2 C 222kPa 5:48  
136.3 C 222kPa 6:18  
136.2 C 224kPa 6:48  
136.3 C 224kPa 7:18  
136.3 C 224kPa 7:18  
VENTING 7:18  
AIR DRYING 8:10  
CYCLE COMPLETE 8:25  
-----

Digital Signature #  
BFCDB6F9D6BCDD10  
-----