
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004936
15:30 09/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:55
STERILIZING 3:49
136.1 C 221kPa 3:49
136.1 C 222kPa 4:19
136.4 C 225kPa 4:49
136.3 C 223kPa 5:19
136.4 C 223kPa 5:49
136.3 C 223kPa 6:19
136.4 C 223kPa 6:49
136.4 C 223kPa 7:19
136.4 C 223kPa 7:19
VENTING 7:19
AIR DRYING 8:17
CYCLE COMPLETE 11:16

Digital Signature #
1A215600640115C3
