
Midmark M11 - v1.0.0
Total Cycles: 1958

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2022 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 168.9 F 3.5
4:00 192.9 F 8.1
6:00 217.0 F 9.7
8:00 229.9 F 8.1
10:00 239.6 F 13.3
12:00 239.2 F 10.6
14:00 236.0 F 8.6
16:00 250.7 F 15.7
18:00 262.7 F 22.6
20:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.0
2:00 271.4 F 28.4
3:00 272.7 F 29.1
4:00 271.5 F 28.4

Min 271.4 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 236.2 F 0.0
10:00 242.0 F 0.0
15:00 228.0 F 0.0
20:00 228.8 F 0.0
25:00 229.7 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

05/26/2022 01:10 PM
ACCEPTED: X
REJECTED:
Operator: U02