
Midmark M11 - v1.0.5
Total Cycles: 3836

Name: RITTER2
Sterilizer ID: V2313228
08/07/2024 04:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.1 F 0.0
2:00 147.2 F 2.1
4:00 156.4 F 5.0
6:00 184.8 F 8.7
8:00 206.2 F 13.1
10:00 218.5 F 6.6
12:00 225.8 F 11.0
14:00 229.8 F 7.4
16:00 241.5 F 12.8
18:00 241.1 F 11.4
20:00 252.8 F 17.4
22:00 263.2 F 23.6
24:00 272.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 271.8 F 29.4
2:00 271.9 F 29.8
3:00 271.2 F 29.1
4:00 272.5 F 30.4

Min 271.0 F 29.0
Max 272.7 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 206.6 F 0.0
10:00 205.2 F 0.0
15:00 199.9 F 0.0
20:00 193.1 F 0.0
25:00 187.5 F 0.0
30:00 184.1 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

08/07/2024 05:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____