
Midmark M11 - v1.0.5
Total Cycles: 881

Name: RITTER2
Sterilizer ID: V2313228
03/16/2022 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 80.6 F 0.0
4:00 96.8 F 0.0
6:00 136.5 F 0.0
8:00 145.2 F 0.3
10:00 175.2 F 3.0
12:00 221.8 F 6.4
14:00 227.8 F 10.4
16:00 229.2 F 8.5
18:00 238.0 F 10.0
20:00 234.1 F 9.2
22:00 243.0 F 12.1
24:00 239.6 F 10.2
26:00 251.8 F 16.2
28:00 262.3 F 22.3
30:00 271.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 271.7 F 29.3
2:00 271.4 F 29.3
3:00 271.4 F 29.2
4:00 271.8 F 29.1

Min 271.4 F 29.1
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 223.5 F 0.0
10:00 232.7 F 0.0
15:00 225.1 F 0.0
20:00 226.8 F 0.0
25:00 230.9 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 31:11
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:09:13
CYCLE COMPLETE

03/16/2022 09:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____