

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1745

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2022 08:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 141.2 F 0.0  
2:00 141.5 F 0.0  
4:00 164.9 F 1.4  
6:00 178.5 F 4.3  
8:00 206.7 F 8.4  
10:00 226.9 F 12.9  
12:00 227.8 F 7.5  
14:00 239.3 F 12.6  
16:00 236.8 F 9.7  
18:00 236.9 F 8.7  
20:00 247.8 F 14.3  
22:00 259.8 F 21.0  
24:00 270.9 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.6 F 29.5  
2:00 272.5 F 28.9  
3:00 271.0 F 28.4  
4:00 272.7 F 29.6

Min 271.0 F 28.4  
Max 273.1 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.9 F 0.0  
5:00 210.5 F 0.0  
10:00 198.7 F 0.0  
15:00 203.0 F 0.0  
20:00 219.5 F 0.0  
25:00 225.5 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 24:49  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:02:51  
CYCLE COMPLETE

03/25/2022 10:32 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02