
Midmark M11 - v1.0.0
Total Cycles: 2550

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2022 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.6 F 0.0
2:00 161.5 F 1.4
4:00 165.4 F 4.9
6:00 190.1 F 9.5
8:00 219.0 F 7.4
10:00 225.2 F 8.8
12:00 235.6 F 11.3
14:00 238.7 F 10.9
16:00 236.2 F 8.7
18:00 250.7 F 15.6
20:00 262.6 F 22.6
22:00 272.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.6 F 29.1
2:00 271.5 F 28.3
3:00 272.2 F 29.1
4:00 272.3 F 28.8

Min 271.2 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 224.7 F 0.0
10:00 237.1 F 0.0
15:00 229.4 F 0.0
20:00 224.5 F 0.0
25:00 230.3 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 22:13
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

11/14/2022 12:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____