

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3232

Name: RITTER 1  
Sterilizer ID: V2178198  
05/15/2023 12:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.6 F 0.0  
2:00 174.0 F 0.0  
4:00 186.9 F 4.0  
6:00 208.8 F 8.8  
8:00 225.4 F 10.1  
10:00 232.5 F 9.7  
12:00 234.9 F 8.5  
14:00 245.2 F 13.5  
16:00 244.3 F 12.4  
18:00 257.3 F 19.2  
20:00 268.6 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 271.7 F 28.7  
2:00 272.5 F 29.0  
3:00 271.8 F 28.4  
4:00 271.4 F 28.5  
  
Min 271.2 F 28.1  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
5:00 212.7 F 0.0  
10:00 230.9 F 0.0  
15:00 229.1 F 0.0  
20:00 229.9 F 0.0  
25:00 231.0 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 21:05  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:59:17  
CYCLE COMPLETE

05/15/2023 01:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_