
Midmark M11 - v1.0.0
Total Cycles: 2479

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2022 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 182.1 F 3.9
4:00 189.7 F 8.7
6:00 210.5 F 10.3
8:00 222.6 F 7.2
10:00 236.7 F 12.3
12:00 236.4 F 9.3
14:00 238.5 F 9.4
16:00 247.5 F 14.0
18:00 259.3 F 20.5
20:00 270.2 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.7 F 28.3
3:00 272.7 F 29.2
4:00 272.2 F 28.7

Min 271.5 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 218.2 F 0.0
10:00 239.6 F 0.0
15:00 228.6 F 0.0
20:00 225.8 F 0.0
25:00 226.8 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:54
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

10/24/2022 02:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____