
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002610
9:41 26/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:43
STERILIZING 3:46
136.5 C 223kPa 3:46
136.4 C 223kPa 4:16
136.4 C 224kPa 4:46
136.1 C 221kPa 5:16
136.3 C 220kPa 5:46
136.1 C 222kPa 6:16
136.4 C 225kPa 6:46
136.1 C 221kPa 7:16
136.1 C 221kPa 7:16
VENTING 7:16
AIR DRYING 8:08
CYCLE COMPLETE 38:08

Digital Signature #
C9798E93CBFB4285
