
Midmark M11 - v1.0.0
Total Cycles: 1065

Name: RITTER 1
Sterilizer ID: V2178198
07/28/2021 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.0 F 0.0
2:00 166.5 F 2.2
4:00 175.8 F 5.5
6:00 207.8 F 10.2
8:00 223.0 F 4.8
10:00 231.3 F 10.0
12:00 233.8 F 7.9
14:00 242.5 F 12.5
16:00 240.8 F 10.7
18:00 253.5 F 17.1
20:00 264.6 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 29.1
2:00 271.6 F 28.4
3:00 271.8 F 28.9
4:00 272.4 F 29.0

Min 271.1 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.8 F 0.0
5:00 228.5 F 0.0
10:00 226.7 F 0.0
15:00 211.8 F 0.0
20:00 203.4 F 0.0
25:00 209.2 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

07/28/2021 11:39 AM
ACCEPTED: X
REJECTED:
Operator: U01