
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006234
11:25 05/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:01
STERILIZING 2:54
136.4 C 220kPa 2:54
136.1 C 222kPa 3:24
136.2 C 223kPa 3:54
136.4 C 224kPa 4:24
136.2 C 222kPa 4:54
136.2 C 222kPa 5:24
136.4 C 221kPa 5:54
136.2 C 222kPa 6:24
136.2 C 222kPa 6:24
VENTING 6:24
AIR DRYING 7:13
CYCLE COMPLETE 7:22

Digital Signature #
55633E0D01A4A3CE
