
Midmark M11 - v1.0.5
Total Cycles: 219

Name: RITTER2
Sterilizer ID: V2313228
07/15/2021 03:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.5 F 0.0
2:00 189.7 F 4.1
4:00 209.8 F 9.8
6:00 227.9 F 7.2
8:00 236.9 F 12.6
10:00 240.3 F 11.3
12:00 240.2 F 10.9
14:00 254.4 F 18.1
16:00 266.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.1 F 29.2
2:00 272.0 F 29.6
3:00 271.8 F 28.9
4:00 272.3 F 29.7

Min 271.1 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
Dry Cycle Aborted

FILLING: 2:51
HEATING: 17:27
STERILIZING: 4:00
VENTING: 3:13
DRYING: 3:53
TOTAL CYCLE: 00:31:25

07/15/2021 04:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____