

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4424

Name: RITTER 1  
Sterilizer ID: V2178198  
04/03/2024 12:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 101.4 F 0.0  
2:00 118.0 F 0.9  
4:00 142.0 F 2.9  
6:00 148.0 F 5.8  
8:00 176.9 F 9.5  
10:00 195.3 F 7.1  
12:00 211.2 F 4.7  
14:00 225.4 F 8.2  
16:00 235.0 F 12.1  
18:00 231.8 F 8.0  
20:00 243.2 F 13.0  
22:00 240.7 F 11.0  
24:00 252.6 F 16.8  
26:00 263.0 F 22.9  
28:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 29.0  
2:00 271.5 F 28.6  
3:00 272.5 F 29.1  
4:00 271.5 F 28.4

Min 271.4 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 206.7 F 0.0  
10:00 199.6 F 0.0  
15:00 193.2 F 0.0  
20:00 187.4 F 0.0  
25:00 192.4 F 0.0  
30:00 193.1 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 28:22  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:06:50  
CYCLE COMPLETE

04/03/2024 01:36 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_