
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000097
9:22 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:54
STERILIZING 5:26
136.0 C 222kPa 5:26
136.2 C 224kPa 5:56
136.3 C 226kPa 6:26
136.2 C 225kPa 6:56
136.2 C 224kPa 7:26
136.3 C 224kPa 7:56
136.3 C 225kPa 8:26
136.3 C 226kPa 8:56
136.3 C 226kPa 8:56
VENTING 8:56
AIR DRYING 10:09
CYCLE COMPLETE 25:09

Digital Signature #
444E2C37FFF32544
