
Midmark M11 - v1.0.5
Total Cycles: 199

Name: RITTER2
Sterilizer ID: V2313228
07/13/2021 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.6 F 0.0
2:00 201.5 F 3.2
4:00 197.0 F 8.2
6:00 214.1 F 10.8
8:00 229.1 F 9.2
10:00 236.2 F 10.0
12:00 244.1 F 13.4
14:00 244.7 F 13.0
16:00 257.6 F 19.9
18:00 268.9 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.3
2:00 272.1 F 29.5
3:00 272.2 F 29.1
4:00 272.3 F 29.7

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 220.2 F 0.0
10:00 232.5 F 0.0
15:00 229.3 F 0.0
20:00 220.3 F 0.0
25:00 216.0 F 0.0
Dry Cycle Aborted

FILLING: 2:01
HEATING: 19:06
STERILIZING: 4:00
VENTING: 3:07
DRYING: 27:18
TOTAL CYCLE: 00:55:31

07/13/2021 10:16 AM
ACCEPTED: X
REJECTED:
Operator: U01