
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003855
8:34 22/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:05
136.4 C 221kPa 3:05
136.1 C 221kPa 3:35
136.4 C 222kPa 4:05
136.4 C 219kPa 4:35
136.1 C 222kPa 5:05
135.8 C 219kPa 5:35
136.4 C 221kPa 6:05
136.4 C 221kPa 6:35
136.4 C 221kPa 6:35
VENTING 6:36
AIR DRYING 7:23
CYCLE COMPLETE 9:40

Digital Signature #
E0A64A2539AC03A3
