
Midmark M11 - v1.0.0
Total Cycles: 3025

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2023 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 182.7 F 3.7
4:00 192.8 F 8.1
6:00 220.3 F 10.2
8:00 228.8 F 8.9
10:00 237.0 F 9.7
12:00 242.1 F 12.0
14:00 240.8 F 10.6
16:00 254.6 F 17.6
18:00 266.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 28.8
2:00 271.2 F 28.2
3:00 272.4 F 29.0
4:00 271.7 F 28.3

Min 271.2 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 219.4 F 0.0
10:00 227.8 F 0.0
15:00 229.4 F 0.0
20:00 228.5 F 0.0
25:00 229.1 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 19:28
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:54
CYCLE COMPLETE

03/22/2023 12:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____