
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001635
8:50 27/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:16
STERILIZING 3:23
136.5 C 223kPa 3:23
136.1 C 221kPa 3:53
136.4 C 223kPa 4:23
136.4 C 220kPa 4:53
136.4 C 223kPa 5:23
136.4 C 224kPa 5:53
136.3 C 223kPa 6:23
136.4 C 222kPa 6:53
136.4 C 222kPa 6:53
VENTING 6:53
AIR DRYING 7:47
CYCLE COMPLETE 7:53

Digital Signature #
9C35B034AFE10282
