
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002205
10:49 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:07
STERILIZING 4:06
136.4 C 226kPa 4:06
136.3 C 225kPa 4:36
136.3 C 225kPa 5:06
136.2 C 224kPa 5:36
136.4 C 226kPa 6:06
136.4 C 223kPa 6:36
136.3 C 225kPa 7:06
136.4 C 226kPa 7:36
136.4 C 226kPa 7:36
VENTING 7:36
AIR DRYING 8:31
CYCLE COMPLETE 10:02

Digital Signature #
C48DA6E909C18A65
