

-----  
Midmark M11 - v1.0.0  
Total Cycles: 932

Name: RITTER 1  
Sterilizer ID: V2178198  
06/23/2021 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.8 F 0.0  
2:00 150.4 F 1.5  
4:00 169.5 F 3.9  
6:00 188.2 F 7.5  
8:00 207.1 F 11.8  
10:00 217.4 F 5.3  
12:00 230.7 F 9.8  
14:00 234.2 F 9.6  
16:00 239.7 F 11.2  
18:00 237.0 F 9.1  
20:00 249.6 F 15.2  
22:00 260.3 F 21.2  
24:00 270.4 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.7 F 29.2  
2:00 271.8 F 28.5  
3:00 271.9 F 28.9  
4:00 272.7 F 29.1

Min 271.4 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 220.9 F 0.0  
10:00 232.0 F 0.0  
15:00 212.0 F 0.0  
20:00 200.9 F 0.0  
25:00 203.8 F 0.0  
30:00 204.3 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 24:54  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:03:11  
CYCLE COMPLETE

06/23/2021 03:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_