
Midmark M11 - v1.0.0
Total Cycles: 2951

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2023 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.0 F 0.0
2:00 89.9 F 0.0
4:00 88.1 F 0.0
6:00 114.5 F 0.2
8:00 145.2 F 1.6
10:00 175.9 F 4.1
12:00 208.3 F 6.9
14:00 206.1 F 10.7
16:00 223.3 F 5.9
18:00 228.3 F 9.0
20:00 239.1 F 12.6
22:00 239.6 F 10.8
24:00 235.6 F 8.6
26:00 249.9 F 15.3
28:00 261.5 F 21.8
30:00 272.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 29.0
2:00 271.3 F 28.4
3:00 272.2 F 29.0
4:00 271.2 F 28.3

Min 271.2 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 206.0 F 0.0
10:00 211.7 F 0.0
15:00 217.6 F 0.0
20:00 220.2 F 0.0
25:00 222.3 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

02/28/2023 09:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____