

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007250  
11:03 23/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:01  
STERILIZING 4:02  
136.0 C 220kPa 4:02  
136.1 C 222kPa 4:32  
136.4 C 224kPa 5:02  
136.4 C 222kPa 5:32  
136.2 C 223kPa 6:02  
136.4 C 225kPa 6:32  
136.4 C 223kPa 7:02  
136.2 C 222kPa 7:32  
136.2 C 222kPa 7:32  
VENTING 7:32  
AIR DRYING 8:27  
CYCLE COMPLETE 23:18

-----  
Digital Signature #  
EE4FBDDDB2256C7C1  
-----