
Midmark M11 - v1.0.0
Total Cycles: 5172

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2024 10:57 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 70.4 F 0.0

C383: HEATUP MODE
DOOR OPEN
CHAMBER TEMP 73F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

12/09/2024 11:00 AM