
Midmark M11 - v1.0.5
Total Cycles: 1979

Name: RITTER2
Sterilizer ID: V2313228
01/30/2023 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.8 F 0.0
2:00 166.9 F 3.6
4:00 189.9 F 7.6
6:00 210.6 F 12.8
8:00 218.9 F 7.3
10:00 231.8 F 12.5
12:00 234.9 F 9.7
14:00 236.4 F 9.0
16:00 248.5 F 15.0
18:00 260.7 F 22.0
20:00 271.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.4 F 29.9
2:00 271.7 F 29.0
3:00 272.6 F 29.8
4:00 271.8 F 29.1

Min 271.5 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 212.0 F 0.0
10:00 223.1 F 0.0
15:00 232.3 F 0.0
20:00 227.7 F 0.0
25:00 225.6 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

01/30/2023 02:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____