
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004616
15:09 25/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:29
STERILIZING 4:33
136.4 C 220kPa 4:33
136.4 C 226kPa 5:03
136.2 C 223kPa 5:33
136.0 C 222kPa 6:03
136.5 C 226kPa 6:33
136.4 C 225kPa 7:03
136.4 C 225kPa 7:33
136.4 C 225kPa 8:03
136.4 C 225kPa 8:03
VENTING 8:03
AIR DRYING 8:54
CYCLE COMPLETE 9:01

Digital Signature #
36677B468C676D22
