
Midmark M11 - v1.0.0
Total Cycles: 5471

Name: RITTER 1
Sterilizer ID: V2178198
04/02/2025 03:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.1 F 0.0
2:00 142.6 F 2.2
4:00 151.7 F 5.2
6:00 181.6 F 8.4
8:00 203.9 F 12.7
10:00 213.4 F 4.4
12:00 222.4 F 8.3
14:00 232.9 F 12.2
16:00 233.3 F 7.9
18:00 243.2 F 12.9
20:00 241.4 F 10.8
22:00 252.4 F 16.4
24:00 262.2 F 22.1
26:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.6 F 29.0
2:00 271.8 F 28.9
3:00 272.3 F 28.9
4:00 271.8 F 29.0

Min 271.6 F 28.6
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 202.3 F 0.0
10:00 212.4 F 0.0
15:00 212.3 F 0.0
20:00 214.4 F 0.0
25:00 216.0 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 27:31
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:06:26
CYCLE COMPLETE

04/02/2025 04:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____