

-----  
Midmark M11 - v1.0.5  
Total Cycles: 541

Name: RITTER2  
Sterilizer ID: V2313228  
11/11/2021 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.2 F 0.0  
2:00 168.9 F 3.0  
4:00 192.6 F 7.8  
6:00 214.8 F 13.5  
8:00 230.2 F 8.7  
10:00 236.9 F 10.6  
12:00 242.9 F 12.4  
14:00 242.4 F 11.6  
16:00 255.3 F 18.5  
18:00 267.0 F 25.6  
20:00 272.9 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.7 F 29.5  
1:00 273.0 F 30.0  
2:00 272.1 F 29.9  
3:00 273.0 F 29.6  
4:00 273.1 F 30.2

Min 271.8 F 29.3  
Max 273.3 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
5:00 236.8 F 0.0  
10:00 223.9 F 0.0  
15:00 222.5 F 0.0  
20:00 208.7 F 0.0  
25:00 199.2 F 0.0  
30:00 192.0 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 20:01  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

11/11/2021 10:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_