
Midmark M11 - v1.0.0
Total Cycles: 2183

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2022 07:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 79.2 F 0.0
4:00 84.7 F 0.0
6:00 115.7 F 0.7
8:00 153.5 F 3.1
10:00 175.5 F 6.4
12:00 194.0 F 9.8
14:00 208.8 F 13.3
16:00 216.4 F 5.2
18:00 226.9 F 8.9
20:00 236.8 F 13.0
22:00 234.5 F 9.2
24:00 240.4 F 10.5
26:00 244.2 F 12.4
28:00 255.8 F 18.4
30:00 265.9 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 271.8 F 28.7
2:00 272.4 F 28.9
3:00 271.3 F 28.2
4:00 272.4 F 29.2

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 207.4 F 0.0
10:00 211.0 F 0.0
15:00 223.1 F 0.0
20:00 237.2 F 0.0
25:00 238.2 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:42
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

07/27/2022 08:59 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____