

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3809

Name: RITTER2  
Sterilizer ID: V2313228  
07/23/2024 02:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 83.3 F 0.0  
2:00 84.2 F 0.0  
4:00 90.4 F 0.2  
6:00 134.6 F 0.5  
8:00 164.7 F 3.2  
10:00 189.1 F 5.9  
12:00 196.5 F 9.1  
14:00 206.5 F 12.3  
16:00 215.6 F 5.3  
18:00 225.9 F 8.9  
20:00 235.7 F 13.0  
22:00 235.2 F 9.6  
24:00 240.2 F 10.8  
26:00 244.6 F 12.9  
28:00 254.9 F 18.3  
30:00 264.3 F 24.0  
32:00 272.0 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.5 F 29.7  
2:00 272.4 F 29.5  
3:00 272.3 F 29.9  
4:00 272.1 F 29.4  
  
Min 271.0 F 28.7  
Max 273.2 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.4 F 0.0  
5:00 208.1 F 0.0  
10:00 201.5 F 0.0  
15:00 192.0 F 0.0  
20:00 193.1 F 0.0  
25:00 194.9 F 0.0  
30:00 195.1 F 0.0  
  
Drying Complete

FILLING: 0:54  
HEATING: 32:15  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:11:00  
CYCLE COMPLETE

07/23/2024 04:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_