

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2309

Name: RITTER2  
Sterilizer ID: V2313228  
04/27/2023 03:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.2 F 0.0  
2:00 170.8 F 3.3  
4:00 183.6 F 7.5  
6:00 210.9 F 12.7  
8:00 223.9 F 7.8  
10:00 235.9 F 13.4  
12:00 238.9 F 11.3  
14:00 239.0 F 10.3  
16:00 253.0 F 17.2  
18:00 264.7 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 271.5 F 29.0  
2:00 271.5 F 28.9  
3:00 271.8 F 29.0  
4:00 271.8 F 29.3  
  
Min 271.2 F 28.9  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.8 F 0.0  
5:00 213.3 F 0.0  
10:00 201.5 F 0.0  
15:00 202.0 F 0.0  
20:00 199.2 F 0.0  
25:00 196.8 F 0.0  
30:00 194.7 F 0.0  
Drying Complete

FILLING: 1:42  
HEATING: 19:44  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:47  
CYCLE COMPLETE

04/27/2023 04:30 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_