
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000328
15:23 17/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:36
STERILIZING 4:34
136.0 C 220kPa 4:34
136.3 C 225kPa 5:04
136.2 C 221kPa 5:34
136.2 C 221kPa 6:04
136.3 C 220kPa 6:34
136.3 C 225kPa 7:04
136.3 C 224kPa 7:34
136.3 C 221kPa 8:04
136.3 C 221kPa 8:04
VENTING 8:04
AIR DRYING 9:00
CYCLE COMPLETE 9:06

Digital Signature #
3064E12C33A92DB3
