
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000945
12:21 09/12/2020
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 4:46
STERILIZING 5:48
136.4 C 220kPa 5:48
136.3 C 224kPa 6:18
136.3 C 224kPa 6:48
136.2 C 223kPa 7:18
136.4 C 225kPa 7:48
136.4 C 222kPa 8:18
136.4 C 222kPa 8:48
136.2 C 223kPa 9:18
136.1 C 222kPa 9:48
136.4 C 225kPa 10:18
136.2 C 224kPa 10:48
CYCLE ABORTED 10:59
STOP BUTTON PRESSED
NOT STERILE
END TIME 10:59

Digital Signature #
1D1FA3C815DB1511
