

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3046

Name: RITTER2  
Sterilizer ID: V2313228  
10/23/2023 05:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.4 F 0.0  
2:00 118.6 F 1.1  
4:00 148.8 F 4.2  
6:00 179.0 F 8.7  
8:00 201.2 F 9.8  
10:00 217.6 F 7.3  
12:00 231.2 F 12.2  
14:00 233.4 F 9.0  
16:00 238.6 F 10.1  
18:00 246.5 F 13.9  
20:00 259.2 F 20.8  
22:00 270.9 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.7 F 29.4  
2:00 271.7 F 29.2  
3:00 272.4 F 29.2  
4:00 271.6 F 29.2

Min 271.4 F 28.7  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.5 F 0.0  
5:00 184.3 F 0.0  
10:00 208.6 F 0.0  
15:00 211.0 F 0.0  
20:00 210.2 F 0.0  
25:00 210.6 F 0.0  
30:00 209.5 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:43  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:01:02  
CYCLE COMPLETE

10/23/2023 06:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_